

## DECLARATION OF CONFORMITY

Manufacturer's Name: Maxtec  
Address: 2305 South 1070 West  
Salt Lake City, Utah 84119  
USA

European Representative: QNET BV  
Kantstraat 19  
NL-5076 NP Haaren  
The Netherlands

Product: Oxygen Dilutor

Model(s): MaxVenturi, Venturi Muffler (Including Pole Mount Bracket Accessory  
R100P41 and R100P44)

Classification & GMDN: Ila Mixer, Gas Breathing 36327 & 46049

Classification criteria: Clause 3.2 Rule 10 of Annex IX of MDD

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and Standards. All supporting documents are retained under the premises of the manufacturer and the European Representative.

Directives: General application directives: Medical Device Directive, COUNCIL DIRECTIVE  
93/42/EEC of 14 June 1993 per Annex II as amended by 2007/47/EC of 5 September 2007

Notified Body: TÜV SÜD Product Service  
RIDLERSTRASSE 65, D-80339 MUNICH, Germany  
Number 0123

EC Certificate No.: G1 16 10 45041 020

Date CE mark was affixed: 10 October 2007

This declaration is considered valid from September 10, 2019 to December 18, 2021.

Signature:  Date: Sept 10, 2019

Name: Tammy Lavery  
Position: Director of Regulatory and Quality

## Applied Standards

The referenced list of harmonized standards for which documented evidence of compliance can be provided includes:

EN ISO 14971:2012  
EN 62366:2008  
EN 1041:2008  
EN ISO 15223-1:2016  
EN ISO 13485:2016  
EN ISO 10993-1:2009  
IEC 60601-1:2005 (EN- 60601-1:2006/AC:2010)  
IEC 60601-1-2:2014 (EN 60601-1-2:2015)